



Project \_\_\_\_\_  
 AIA # \_\_\_\_\_ SIS # \_\_\_\_\_  
 Item # \_\_\_\_\_ Quantity \_\_\_\_\_ C.S.I. Section 114000



# HEAT DISSIPATION GUIDE

## For CLeN Series

CLEAN SERIES – HEAT DISSIPATION GUIDE

Model	Electric Heat	Steam Heat	Gas Heat	15 kW Electric Booster	30 kW Electric Booster	Steam Booster	Latent Heat (BTU/Hr.)	Sensible Heat (BTU/Hr.)
CL44eN-BAS, CL54eN-BAS	X						14,700	6,300
	X			X			28,300	12,100
	X				X		41,800	17,900
	X					X	30,100	12,900
		X					17,900	7,700
		X		X			31,400	13,500
		X			X		44,900	19,300
		X				X	33,300	14,300
			X				21,800	9,300
			X	X			35,400	15,200
			X		X		48,900	21,000
CL64eN-BAS						X	37,200	15,900
	X						24,800	10,600
	X			X			38,400	16,400
	X				X		51,900	22,200
	X					X	40,200	17,200
		X					35,600	15,300
		X		X			49,200	21,100
		X			X		62,700	26,900
		X				X	51,000	21,900
			X				43,500	18,700
			X	X			57,100	24,500
CLPS66eN-BAS, CLPS76eN-BAS, CLCS66eN-BAS, CLCS76eN-BAS			X				22,900	9,800
			X	X			36,400	15,600
			X		X		50,000	21,400
			X			X	38,300	16,400
	X						15,800	6,800
	X			X			29,300	12,600
	X				X		42,900	18,400
	X					X	31,200	13,400
		X					18,900	8,100
		X		X			32,500	13,900
		X			X		46,000	19,700

**NOTES:** 1. For electric blower dryer add: 11,200 BTU/hr. to latent heat; 4,800 BTU/hr. to sensible heat  
 2. For steam blower dryer add: 16,900 BTU/hr. to latent heat; 7,200 BTU/hr. to sensible heat

Approved by \_\_\_\_\_ Date \_\_\_\_\_ Approved by \_\_\_\_\_ Date \_\_\_\_\_



# HEAT DISSIPATION GUIDE

For CLeN Series

Model	Electric Heat	Steam Heat	Gas Heat	15 kW Electric Booster	30 kW Electric Booster	Steam Booster	Latent Heat (BTU/Hr.)	Sensible Heat (BTU/Hr.)
CLPS86eN-BAS	X						25,900	11,100
	X			X			39,400	16,900
	X				X		53,000	22,700
	X					X	41,300	17,700
		X					36,700	15,700
		X		X			50,200	21,500
		X			X		63,800	27,300
		X				X	52,100	22,300
			X				44,600	19,100
			X	X			58,100	24,900
			X		X		71,700	30,700
		X			X	60,000	25,700	
CL44eN-EGR, CL54eN-EGR	X						14,900	6,400
	X				X		42,000	18,000
	X					X	30,300	13,000
		X					18,000	7,700
		X			X		45,100	19,300
		X				X	33,400	14,300
			X				22,000	9,400
			X		X		49,100	21,000
		X			X	37,400	16,000	
CLPS66eN-EGR, CLPS76eN-EGR, CLCS66eN-EGR, CLCS76eN-EGR	X						16,000	6,800
	X				X		43,100	18,500
	X					X	31,400	13,400
		X					19,100	8,200
		X			X		46,200	19,800
		X				X	34,500	14,800
			X				23,100	9,900
			X		X		50,200	21,500
		X			X	38,500	16,800	
CL64eN-EGR	X						25,000	10,700
	X				X		52,100	22,300
	X					X	40,400	17,300
		X					35,600	15,300
		X			X		62,700	26,900
		X				X	51,000	21,900
			X				43,500	18,700
			X		X		70,600	30,300
		X			X	58,900	25,300	

**NOTES:** 1. For electric blower dryer add: 11,200 BTU/hr. to latent heat; 4,800 BTU/hr. to sensible heat  
 2. For steam blower dryer add: 16,900 BTU/hr. to latent heat; 7,200 BTU/hr. to sensible heat



# HEAT DISSIPATION GUIDE

For CLeN Series

Model	Electric Heat	Steam Heat	Gas Heat	15 kW Electric Booster	30 kW Electric Booster	Steam Booster	Latent Heat (BTU/Hr.)	Sensible Heat (BTU/Hr.)
CLPS86eN-EGR	X						26,100	11,200
	X				X		53,200	22,800
	X					X	41,500	17,800
		X					36,900	15,800
		X			X		64,000	27,400
		X				X	52,300	22,400
			X				44,800	19,200
			X		X		71,900	30,800
			X			X	60,200	25,800
CL44eN-ADV, CL54eN-ADV	X				X		42,200	18,100
CLPS66eN-ADV, CLPS76eN-ADV	X				X		43,200	18,500
CL64eN-ADV	X				X		52,300	22,400
CLPS86eN-ADV	X				X		54,400	23,300
CL44eNVL, CLPS66eNVL	X				X		3,100	6,100

**NOTES:** 1. For electric blower dryer add: 11,200 BTU/hr. to latent heat; 4,800 BTU/hr. to sensible heat  
 2. For steam blower dryer add: 16,900 BTU/hr. to latent heat; 7,200 BTU/hr. to sensible heat